

Certificate of Compliance


This is to certify that

EDGE ANALYTICAL, Inc.,
An Accredited Drinking Water Laboratory,
Certification number 046, has completed the analysis of

PURETAP WATER DISTILLERS LTD

“Spring Water”

on April 7, 2015, according to the CBWA Model Code testing requirements for bottled drinking water. All parameters were found to be in compliance with CBWA's published Standard of Quality limits for bottled water.



Lawrence J. Henderson, PhD
Director of Laboratories

The signature of Lawrence J. Henderson, PhD, is written in black ink above a horizontal line. Below the line, his name and title are printed in a blue, sans-serif font.



| | | | | |
|----------------|----------------------------|--------------------------|-----------------------|-----------------------------|
| Burlington, WA | Corporate Laboratory (a) | 1620 S Walnut St | Burlington, WA 98233 | 800.755.9295 • 360.757.1400 |
| Bellingham, WA | Microbiology (b) | 805 Orchard Dr Ste 4 | Bellingham, WA 98225 | 360.715.1212 |
| Portland, OR | Microbiology/Chemistry (c) | 9150 SW Pioneer Ct Ste W | Wilsonville, OR 97070 | 503.682.7802 |
| Corvallis, OR | Microbiology (d) | 540 SW Third Street | Corvallis, OR 97333 | 541.753.4946 |

April 7, 2015

Page 1 of 1

Mr. Mario Iorio
Puretap Water Distillers Ltd
950 Verbena Road
Mississauga, ON L5T 1T6
RE: 15-04750 - Spring Water Product Testing

Dear Mr. Mario Iorio,

Your project: Spring Water Product Testing, was received on Friday March 13, 2015.

All samples were analyzed within the accepted holding times, were appropriately preserved and were analyzed according to approved analytical protocols. The quality control data was within laboratory acceptance limits, unless specified in the QA reports.

If you have questions phone us at 800 755-9295.

Respectfully

Patrick Miller, MS
QA Officer

Enclosures: Data Report



| | | | | |
|----------------|----------------------------|--------------------------|-----------------------|-----------------------------|
| Burlington, WA | Corporate Laboratory (a) | 1620 S Walnut St | Burlington, WA 98233 | 800.755.9295 • 360.757.1400 |
| Bellingham, WA | Microbiology (b) | 805 Orchard Dr Ste 4 | Bellingham, WA 98225 | 360.715.1212 |
| Portland, OR | Microbiology/Chemistry (c) | 9150 SW Pioneer Ct Ste W | Wilsonville, OR 97070 | 503.682.7802 |
| Corvallis, OR | Microbiology (d) | 540 SW Third Street | Corvallis, OR 97333 | 541.753.4946 |



BOTTLED WATER STANDARD OF QUALITY REPORT

Client Name: Puretap Water Distillers Ltd
 950 Verbena Road
 Mississauga, ON L5T 1T6

Reference Number: **15-04750**

Authorized by:

Pat Miller
 Patrick Miller, MS
 QA Officer

Project: Spring Water Product Testing

Field ID: 15-061

Sample Description: Spring Water

Sampled By:

Sample Date: 03/16/2015

Lab Number: 10870

Report Date: 04/07/2015

Approved By: bj,co,fm,mcs,mvp,pdm,pms

Inorganic Chemicals (IOCs)

| CAS ID# | COMPOUNDS | RESULT | SOQ | MRL | Units | Method | Lab | COMMENT |
|------------|-----------------------|--------|-------|--------|-------|--------|-----|---------|
| 7440-36-0 | ANTIMONY | ND | 0.006 | 0.001 | mg/L | 200.8 | a | |
| 7440-38-2 | ARSENIC | ND | 0.010 | 0.001 | mg/L | 200.8 | a | |
| 7440-39-3 | BARIUM | 0.060 | 1.0 | 0.001 | mg/L | 200.8 | a | |
| 7440-41-7 | BERYLLIUM | ND | 0.004 | 0.001 | mg/L | 200.8 | a | |
| 7440-43-9 | CADMIUM | ND | 0.005 | 0.001 | mg/L | 200.8 | a | |
| 7440-47-3 | CHROMIUM | 0.002 | 0.05 | 0.001 | mg/L | 200.8 | a | |
| 16984-48-8 | FLUORIDE | ND | 1.0 | 0.10 | mg/L | 300.0 | a | |
| 7439-92-1 | LEAD | ND | 0.005 | 0.001 | mg/L | 200.8 | a | |
| 7439-97-6 | MERCURY | ND | 0.001 | 0.0002 | mg/L | 245.1 | a | |
| 7440-02-0 | NICKEL | 0.001 | 0.1 | 0.001 | mg/L | 200.8 | a | |
| 14797-55-8 | NITRATE-N | 0.86 | 10 | 0.10 | mg/L | 300.0 | a | |
| 14797-65-0 | NITRITE-N | ND | 1.0 | 0.10 | mg/L | 300.0 | a | |
| E-10128 | TOTAL NITRATE/NITRITE | 0.86 | 10 | 0.10 | mg/L | 300.0 | a | |
| 7782-49-2 | SELENIUM | ND | 0.010 | 0.005 | mg/L | 200.8 | a | |
| 7440-28-0 | THALLIUM | ND | 0.002 | 0.001 | mg/L | 200.8 | a | |

Notation:

A Result of "ND" indicates that the compound was not detected above the Lab's Reporting Limit - MRL.

SOQ - Standard of Quality, maximum permissible level of a contaminant in water established by EPA, NPDWR or IBWA.

MRL - Method Reporting Limit .

An * in front of the parameter name indicates it is not NELAP accredited but it is accredited through WSDOH or USEPA Region 10.

These test results meet all the requirements of NELAC, unless otherwise stated in writing, and relate only to these samples. If you have any questions concerning this report contact us at the above phone number.

BOTTLED WATER STANDARD OF QUALITY REPORT

| CAS ID# | COMPOUNDS | RESULT | SOQ | MRL | Units | Method | Lab | COMMENT |
|------------|------------------------------|--------|-------|-------|-------|-------------|-----|---------|
| 57-12-5 | * CYANIDE, FREE | ND | 0.1 | 0.010 | mg/L | OIA-1677-DW | a | |
| 7429-90-5 | ALUMINUM | ND | 0.2 | 0.010 | mg/L | 200.7 | a | |
| 16887-00-6 | CHLORIDE | 48 | 250 | 0.1 | mg/L | 300.0 | a | |
| 7440-50-8 | COPPER | ND | 1.0 | 0.005 | mg/L | 200.8 | a | |
| 7439-89-6 | IRON | ND | 0.3 | 0.050 | mg/L | 200.7 | a | |
| 7439-96-5 | MANGANESE | ND | 0.05 | 0.001 | mg/L | 200.8 | a | |
| 7440-22-4 | SILVER | ND | 0.025 | 0.010 | mg/L | 200.8 | a | |
| 14808-79-8 | SULFATE | 20 | 250 | 0.2 | mg/L | 300.0 | a | |
| E-10173 | TOTAL DISSOLVED SOLIDS (TDS) | 334 | 500 | 10 | mg/L | SM2540 C | a | |
| 7440-66-6 | ZINC | ND | 5.0 | 0.005 | mg/L | 200.8 | a | |

Notation:

A Result of "ND" indicates that the compound was not detected above the Lab's Reporting Limit - MRL.
 SOQ - Standard of Quality, maximum permissible level of a contaminant in water established by EPA, NPDWR or IBWA.
 MRL - Method Reporting Limit .

An * in front of the parameter name indicates it is not NELAP accredited but it is accredited through WSDOH or USEPA Region 10.

These test results meet all the requirements of NELAC, unless otherwise stated in writing, and relate only to these samples.

BOTTLED WATER STANDARD OF QUALITY REPORT

Volatile Organic Chemicals (VOCs)

| CAS ID# | COMPOUNDS | RESULT | SOQ | MRL | Units | Method | Lab | COMMENT |
|-----------|--------------------------------------|--------|--------|--------|-------|--------|-----|---------|
| 71-55-6 | 1,1,1 - TRICHLOROETHANE | ND | 0.030 | 0.0004 | mg/L | 524.2 | a | |
| 79-00-5 | 1,1,2 - TRICHLOROETHANE | ND | 0.003 | 0.0004 | mg/L | 524.2 | a | |
| 75-35-4 | 1,1 - DICHLOROETHYLENE | ND | 0.002 | 0.0004 | mg/L | 524.2 | a | |
| 120-82-1 | 1,2,4 - TRICHLOROBENZENE | ND | 0.009 | 0.0004 | mg/L | 524.2 | a | |
| 107-06-2 | 1,2 - DICHLOROETHANE | ND | 0.002 | 0.0004 | mg/L | 524.2 | a | |
| 78-87-5 | 1,2 - DICHLOROPROPANE | ND | 0.005 | 0.0004 | mg/L | 524.2 | a | |
| 71-43-2 | BENZENE | ND | 0.001 | 0.0004 | mg/L | 524.2 | a | |
| 56-23-5 | CARBON TETRACHLORIDE | ND | 0.002 | 0.0004 | mg/L | 524.2 | a | |
| 156-59-2 | CIS - 1,2 - DICHLOROETHYLENE | ND | 0.070 | 0.0004 | mg/L | 524.2 | a | |
| 156-60-5 | TRANS - 1,2 - DICHLOROETHYLENE | ND | 0.100 | 0.0004 | mg/L | 524.2 | a | |
| 100-41-4 | ETHYLBENZENE | ND | 0.0024 | 0.0004 | mg/L | 524.2 | a | |
| 75-09-2 | METHYLENE CHLORIDE (DICHLOROMETHANE) | ND | 0.003 | 0.0004 | mg/L | 524.2 | a | |
| 108-90-7 | MONOCHLOROBENZENE | ND | 0.050 | 0.0004 | mg/L | 524.2 | a | |
| 95-50-1 | O - DICHLOROBENZENE | ND | 0.200 | 0.0004 | mg/L | 524.2 | a | |
| 106-46-7 | P - DICHLOROBENZENE | ND | 0.005 | 0.0004 | mg/L | 524.2 | a | |
| 100-42-5 | STYRENE | ND | 0.100 | 0.0004 | mg/L | 524.2 | a | |
| 127-18-4 | TETRACHLOROETHYLENE | ND | 0.001 | 0.0004 | mg/L | 524.2 | a | |
| 108-88-3 | TOLUENE | ND | 0.024 | 0.0004 | mg/L | 524.2 | a | |
| 79-01-6 | TRICHLOROETHYLENE | ND | 0.001 | 0.0004 | mg/L | 524.2 | a | |
| 75-01-4 | VINYL CHLORIDE | ND | 0.002 | 0.0004 | mg/L | 524.2 | a | |
| 1330-20-7 | XYLENES (TOTAL) | ND | 0.300 | 0.0004 | mg/L | 524.2 | a | |
| 75-27-4 | BROMODICHLOROMETHANE | ND | | 0.0004 | mg/L | 524.2 | a | |
| 124-48-1 | CHLORODIBROMOMETHANE | ND | | 0.0004 | mg/L | 524.2 | a | |
| 67-66-3 | CHLOROFORM | ND | | 0.0004 | mg/L | 524.2 | a | |
| 75-25-2 | BROMOFORM | ND | | 0.0004 | mg/L | 524.2 | a | |
| E-14471 | TOTAL TRIHALOMETHANE | ND | 0.010 | 0.0004 | mg/L | 524.2 | a | |
| 1634-04-4 | METHYL TERT-BUTYL ETHER | ND | 0.070 | 0.0004 | mg/L | 524.2 | a | |

Notation:

A Result of "ND" indicates that the compound was not detected above the Lab's Reporting Limit - MRL.
 SOQ - Standard of Quality, maximum permissible level of a contaminant in water established by EPA, NPDR or IBWA.
 MRL - Method Reporting Limit .

An * in front of the parameter name indicates it is not NELAP accredited but it is accredited through WSDOH or USEPA Region 10.

These test results meet all the requirements of NELAC, unless otherwise stated in writing, and relate only to these samples.

BOTTLED WATER STANDARD OF QUALITY REPORT

| CAS ID# | COMPOUNDS | RESULT | SOQ | MRL | Units | Method | Lab | COMMENT |
|------------|-----------------------------------|--------|---------|---------|-------|--------|-----|----------------------|
| 80-05-7 | * BISPHENOL-A | 1.5 | | 1 | ug/L | 525.2 | a | qualitative analysis |
| 93-72-1 | 2,4,5 - TP (SILVEX) | ND | 0.010 | 0.0002 | mg/L | 515.4 | a | |
| 94-75-7 | 2,4 - D | ND | 0.070 | 0.0001 | mg/L | 515.4 | a | |
| 15972-60-8 | ALACHLOR | ND | 0.002 | 0.0002 | mg/L | 525.2 | a | |
| 116-06-3 | ALDICARB | ND | 0.003 | 0.001 | mg/L | 531.2 | a | |
| 1646-88-4 | ALDICARB SULFONE | ND | 0.003 | 0.001 | mg/L | 531.2 | a | |
| 1646-87-3 | ALDICARB SULFOXIDE | ND | 0.004 | 0.001 | mg/L | 531.2 | a | |
| 1912-24-9 | ATRAZINE | ND | 0.003 | 0.0001 | mg/L | 525.2 | a | |
| 1563-66-2 | CARBOFURAN | ND | 0.040 | 0.0009 | mg/L | 531.2 | a | |
| 57-74-9 | CHLORDANE | ND | 0.0005 | 0.0002 | mg/L | 508.1 | a | |
| 75-99-0 | DALAPON | ND | 0.200 | 0.001 | mg/L | 515.4 | a | |
| 96-12-8 | DIBROMOCHLOROPROPANE (DBCP) | ND | 0.0002 | 0.00002 | mg/L | 504.1 | a | |
| 88-85-7 | DINOSEB | ND | 0.007 | 0.0002 | mg/L | 515.4 | a | |
| 72-20-8 | ENDRIN | ND | 0.002 | 0.00001 | mg/L | 525.2 | a | |
| 106-93-4 | 1,2 - DIBROMOETHANE (EDB) | ND | 0.00005 | 0.00001 | mg/L | 504.1 | a | |
| 76-44-8 | HEPTACHLOR | ND | 0.0004 | 0.00004 | mg/L | 525.2 | a | |
| 1024-57-3 | HEPTACHLOR EPOXIDE "B" | ND | 0.0002 | 0.00002 | mg/L | 525.2 | a | |
| 58-89-9 | LINDANE (BHC - GAMMA) | ND | 0.0002 | 0.00002 | mg/L | 525.2 | a | |
| 72-43-5 | METHOXYCHLOR | ND | 0.040 | 0.0001 | mg/L | 525.2 | a | |
| 23135-22-0 | OXAMYL (VYDATE) | ND | 0.200 | 0.002 | mg/L | 531.2 | a | |
| 87-86-5 | PENTACHLOROPHENOL | ND | 0.001 | 0.00004 | mg/L | 515.4 | a | |
| 1918-02-1 | PICLORAM | ND | 0.190 | 0.0001 | mg/L | 515.4 | a | |
| 1336-36-3 | POLYCHLORINATED BIPHENYLS (PCB) | ND | 0.0005 | 0.0001 | mg/L | 508.1 | a | |
| 122-34-9 | SIMAZINE | ND | 0.004 | 0.00007 | mg/L | 525.2 | a | |
| 8001-35-2 | TOXAPHENE | ND | 0.003 | 0.001 | mg/L | 508.1 | a | |
| 41903-57-5 | DIOXIN (2,3,7,8-TETRACHLORODIBEN; | ND | 3E-8 | 5E-9 | mg/L | 1613 | | Analyzed by PACE_MN |
| 85-00-7 | DIQUAT | ND | 0.020 | 0.0004 | mg/L | 549.2 | a | |
| 145-73-3 | ENDOTHALL | ND | 0.100 | 0.009 | mg/L | 548.1 | a | |
| 1071-83-6 | GLYPHOSATE | ND | 0.280 | 0.006 | mg/L | 547 | a | |
| 50-32-8 | BENZO(A)PYRENE | ND | 0.0002 | 0.00002 | mg/L | 525.2 | a | |
| 103-23-1 | DI(2-ETHYLHEXYL)-ADIPATE | ND | 0.400 | 0.0006 | mg/L | 525.2 | a | |
| 117-81-7 | DI(2-ETHYLHEXYL)-PHTHALATE | ND | 0.006 | 0.0006 | mg/L | 525.2 | a | |
| 118-74-1 | HEXACHLOROBENZENE | ND | 0.001 | 0.0001 | mg/L | 525.2 | a | |
| 77-47-4 | HEXACHLOROCYCLO-PENTADIENE | ND | 0.050 | 0.0001 | mg/L | 525.2 | a | |
| E-10253 | * PHENOLICS | ND | 0.001 | 0.001 | mg/L | 420.4 | | Analyzed by NSF |

Notation:

A Result of "ND" indicates that the compound was not detected above the Lab's Reporting Limit - MRL.

SOQ - Standard of Quality, maximum permissible level of a contaminant in water established by EPA, NPDWR or IBWA.

MRL - Method Reporting Limit .

An * in front of the parameter name indicates it is not NELAP accredited but it is accredited through WSDOH or USEPA Region 10.

These test results meet all the requirements of NELAC, unless otherwise stated in writing, and relate only to these samples.

BOTTLED WATER STANDARD OF QUALITY REPORT

Water Properties

| CAS ID# | COMPOUNDS | RESULT | SOQ | MRL | Units | Method | Lab | COMMENT |
|---------|-------------------|-------------|---------|------|-------------|----------|-----|--------------------|
| E-11712 | COLOUR | ND | 5 | 5 | Color Units | SM2120 B | a | pH:7.54 |
| E-10617 | TURBIDITY | ND | 0.5 | 0.10 | NTU | 180.1 | a | |
| E-10139 | HYDROGEN ION (pH) | 7.54 | 5.0-8.5 | | pH Units | 150.1 | a | |
| E-11734 | * ODOUR | 1.0 | 3 | 1 | TON | SM2150 | a | Temperature:38.4 C |

Notation:

A Result of "ND" indicates that the compound was not detected above the Lab's Reporting Limit - MRL.

SOQ - Standard of Quality, maximum permissible level of a contaminant in water established by EPA, NPDWR or IBWA.

MRL - Method Reporting Limit .

An * in front of the parameter name indicates it is not NELAP accredited but it is accredited through WSDOH or USEPA Region 10.

These test results meet all the requirements of NELAC, unless otherwise stated in writing, and relate only to these samples.

BOTTLED WATER STANDARD OF QUALITY REPORT

Disinfectants/DBP

| CAS ID# | COMPOUNDS | RESULT | SOQ | MRL | Units | Method | Lab | COMMENT |
|------------|--------------------------|---------------|-------|-------|-------|---------------|-----|---------|
| NA | * HAA(5) | ND | 0.060 | 0.001 | mg/L | 552.3 | a | |
| 15541-45-4 | BROMATE | 0.0082 | 0.010 | 0.001 | mg/L | 300.1 | a | |
| 10049-04-4 | * CHLORINE DIOXIDE | ND | 0.8 | 0.10 | mg/L | SM4500-CIO2 I | a | |
| 7758-19-2 | * CHLORITE | ND | 1.00 | 0.010 | mg/L | 300.1 | a | |
| NA | * CHLOROAMINES TOTAL | ND | 3.0 | 0.05 | mg/L | SM4500-CI G | a | |
| 7782-50-5 | * FREE CHLORINE RESIDUAL | ND | 0.1 | 0.05 | mg/L | SM4500-CI G | a | |

Notation:

A Result of "ND" indicates that the compound was not detected above the Lab's Reporting Limit - MRL.

SOQ - Standard of Quality, maximum permissible level of a contaminant in water established by EPA, NPDWR or IBWA.

MRL - Method Reporting Limit .

An * in front of the parameter name indicates it is not NELAP accredited but it is accredited through WSDOH or USEPA Region 10.

These test results meet all the requirements of NELAC, unless otherwise stated in writing, and relate only to these samples.

BOTTLED WATER STANDARD OF QUALITY REPORT

Radiological Contaminants

| CAS ID# | COMPOUNDS | RESULT | SOQ | MRL | Units | Method | Lab | COMMENT |
|------------|-------------|--------|-----|-----|-------|--------|-----|------------------|
| 12587-46-1 | GROSS ALPHA | ND | 15 | 3 | pCi/L | 900.0 | | Analyzed by Pace |
| 12587-47-2 | GROSS BETA | ND | 50 | 4 | pCi/L | 900.0 | | Analyzed by Pace |

Notation:

A Result of "ND" indicates that the compound was not detected above the Lab's Reporting Limit - MRL.

SOQ - Standard of Quality, maximum permissible level of a contaminant in water established by EPA, NPDWR or IBWA.

MRL - Method Reporting Limit .

An * in front of the parameter name indicates it is not NELAP accredited but it is accredited through WSDOH or USEPA Region 10.

These test results meet all the requirements of NELAC, unless otherwise stated in writing, and relate only to these samples.



| | | | | |
|----------------|----------------------------|--------------------------|-----------------------|-----------------------------|
| Burlington, WA | Corporate Laboratory (a) | 1620 S Walnut St | Burlington, WA 98233 | 800.755.9295 • 360.757.1400 |
| Bellingham, WA | Microbiology (b) | 805 Orchard Dr Ste 4 | Bellingham, WA 98225 | 360.715.1212 |
| Portland, OR | Microbiology/Chemistry (c) | 9150 SW Pioneer Ct Ste W | Wilsonville, OR 97070 | 503.682.7802 |
| Corvallis, OR | Microbiology (d) | 540 SW Third Street | Corvallis, OR 97333 | 541.753.4946 |



Data Report

Client Name: Puretap Water Distillers Ltd
950 Verbena Road
Mississauga, ON L5T 1T6

Reference Number: **15-04750**
Project: Spring Water Product Testing

Report Date: 4/7/15

Date Received: 3/13/15

Approved by:.mvp,sps

Authorized by:


Patrick Miller, MS
QA Officer

Sample Description: 15-061 - Spring Water
Lab Number: 10870 Sample Comment:
Sample Date: 3/16/15 9:30 am
Collected By:

| CAS ID# | Parameter | Result | PQL | MDL | Units | DF | Method | Lab | Analyzed | Analyst | Batch | Comment |
|-----------|--------------------------------------|--------|------|-------|-------|-----|----------|-----|----------|---------|---------------|---------|
| 7440-70-2 | CALCIUM | 105.7 | 1 | 0.009 | mg/L | 1.0 | 200.7 | a | 3/19/15 | BJ | 200.7-150319B | |
| E-11778 | HARDNESS as Calcium Carbonate | 328.6 | 10 | 0.01 | mg/L | 1.0 | 200.7 | a | 3/19/15 | BJ | 200.7-150319B | |
| 7439-95-4 | MAGNESIUM | 15.7 | 1 | 0.001 | mg/L | 1.0 | 200.7 | a | 3/19/15 | BJ | 200.7-150319B | |
| 7440-09-7 | POTASSIUM | 1.3 | 1.0 | 0.1 | mg/L | 1.0 | 200.7 | a | 3/19/15 | BJ | 200.7-150319B | |
| 7440-23-5 | SODIUM | 22.8 | 0.50 | 0.05 | mg/L | 1.0 | 200.7 | a | 3/19/15 | BJ | 200.7-150319B | |
| E-10184 | ELECTRICAL CONDUCTIVITY | 602 | 10 | | uS/cm | 1.0 | SM2510 B | a | 3/16/15 | SRF | EC_150316 | |

Notes:

ND = Not detected above the listed practical quantitation limit (PQL) or not above the Method Detection Limit (MDL), if requested.

PQL = Practical Quantitation Limit is the lowest level that can be achieved within specified limits of precision and accuracy during routine laboratory operating conditions.

D.F. - Dilution Factor

An * in front of the parameter name indicates it is not NELAP accredited but it is accredited through WSDOH or USEPA Region 10.

These test results meet all the requirements of NELAC, unless otherwise stated in writing, and relate only to these samples. Estimates of uncertainty are not included in this report. If this information is required please contact us at the phone number listed in the report header.

If you have any questions concerning this report contact us at the above phone number.